


***BILI-THERAPY** Pad Type*

LED Phototherapy Unit




BILI-THERAPY Pad Type delivers effective phototherapy for the treatment of indirect hyperbilirubinemia from the back of the newborn.

•Relative irradiance



Effective wavelength from the light source of blue LED will encourage newborn's early discharge from the hospital.



BILI-THERAPY Pad Type can be used in a bassinet next to the mother's bed.

Specifications

Atom's code. 60370 BILI-THERAPY Pad Type	
Power requirements	Rating:AC100~240V Power consumption:68VA Frequency:50/60Hz
Classification	Class I, Type BF
Light source	Blue LEDs
Irradiance level	High:53μW/cm ² /nm (3.2mW/cm ²)±25% Medium:40μW/cm ² /nm (2.4mW/cm ²)±25% Low:26.5μW/cm ² /nm (1.6mW/cm ²)±25% *(obtained by the BiliBlanket® Meter II placed on the pad surface without a pad Cover)
Dimensions	15(W)×22(D)×16(H)cm
Weight	Approx. 2.1kg
Accessories	Power cord (AC120V)...1 Power cord (AC230V)...1 Operation Manual1



BILI-THERAPY Pad Type along with BILI-THERAPY Spot Type can effectively provide double phototherapy treatment without interfering with visualisation of the newborn.

Option & Supplies

- Atom's code. 60371 **Pad (L)** For BILI-THERAPY Pad Type
- Atom's code. 60372 **Pad (S)** For BILI-THERAPY Pad Type
- Atom's code. 60373 **Pad Cover (L)**
For BILI-THERAPY Pad Type (50 pcs/box)
- Atom's code. 60374 **Pad Cover (S)**
For BILI-THERAPY Pad Type (50 pcs/box)
- Atom's code. 60375 **BILI-Holder**
For BILI-THERAPY Pad Type/For F-Rail
- Atom's code. 60376 **BILI-Stand** For BILI-THERAPY Pad Type
- Atom's code. 68448 **Phototherapy Analyzer II**

Manufacturer: ATOM MEDICAL CORPORATION
<http://www.atomed.co.jp>

Exporter: ATOM MEDICAL INTERNATIONAL, INC.
Iwakata Bldg., 3-18-16, Hongo, Bunkyo-ku, Tokyo, JAPAN
Phone: + 81 3 3815 2941 Fax: + 81 3 3812 9670/4080
<http://www.atom-ami.co.jp>